

## Dispensing Booth



### Specifications:

MOC	: SS 304,316.
Pre-filter	: Air Micron Make EU 4 rating with suitable size having efficiency 90% down to 10 micron.
Mini-plate Hepa filter	: HEPA Filter having EU 13 or EU14 rating with suitable size having efficiency 99.997% down to 0.3 micron.
Exhaust Filter	: HEPA Filter having EU 13 or EU14 rating with suitable size having efficiency 99.997% down to 0.3 micron.
Air Cleanliness	: Class ISO 5 (as per ISO 14644-1).
Air Flow Rate	: $0.45 \pm 0.1$ m/sec.
Illumination Of Light	: Optimum.
Vibration Level	: Minimum.
Magnehelic Differential Pressure Gauge	: (make : USA) To measure differential pressure across the supply HEPA filter
Motor Blower Assembly	: Statically & Dynamically balanced with suitable rating and size. (Induction motor with centrifugal blower) A Mounting unique spring suspension system to take care of vibration and noise.
Side Panels	: Double skinned PUFF insulated side panels.
3 Pins Switch & Socket	: 15 AMPS - 1 set in each unit.
Feather Touch Switch	: On/Off switches for motor blower & CFL lights.
Working Sizes	: 3 X 2 X 6, 4 X 2 X 6, 6 X 2 X 6 3 X 4 X 6, 4 X 4 X 6, 6 X 4 X 6



**Rombai**  
**INSTRUMENTS**

Url: [www.rombai.com](http://www.rombai.com) • Email: [rombaiinstruments@gmail.com](mailto:rombaiinstruments@gmail.com)